

CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Medical Devices Regulations (CMDR).

Product Name: Xpert® Carba-R

Catalogue Part No.: GXCARBARP-CE-10

Kit Lot No.: 1001447003

Cartridge Lot No.: 11803

Kit Expiration Date: 2026-08-03

<u>Legal Manufacturer</u> <u>Manufacturing Facility</u>

Cepheid 904 Caribbean Drive Sunnyvale, CA 94089 USA Cepheid AB
Röntgenvägen 5
SE-171 54 Solna
Sweden

Solna Sunnyvale

Functional Testing

Test Description	Acceptance Criteria	Test Result
LOW Positive	IMP DETECTED; VIM DETECTED, NDM DETECTED; KPC DETECTED; OXA48 DETECTED	Passed
HIGH Positive	IMP DETECTED; VIM DETECTED, NDM DETECTED; KPC DETECTED; OXA48 DETECTED	Passed
Negative	IMP NOT DETECTED; VIM NOT DETECTED; NDM NOT DETECTED; KPC NOT DETECTED; OXA48 NOT DETECTED	Passed

[If checked this document is produced electronically and valid without a wet signature.

Samaneh Vahid Gharavi (Sep 23, 2024 13:17 GMT+2)	
Signature of Quality Assurance	Date

Name: Samaneh Vahid

Title: Senior QA Analyst