

CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive, the Canadian Medical Devices Regulations (CMDR), and the China National Medical Products Administration's Quality System Requirements.

Product Name: Xpert® Xpress Flu/RSV

Cepheid Catalogue Part No.: XPRSFLU/RSV-CN-10

Kit Lot No.: 1001452780

Cartridge Lot No.: 46507

Kit Expiration Date: 2026-09-27

Legal Manufacturer **Manufacturing Facility**

Cepheid Cepheid

904 Caribbean Drive 904 Caribbean Drive Sunnyvale, CA 94089 Sunnyvale, CA 94089

USA USA

Solna

Sunnyvale

Functional Testing

Test Description	Acceptance Criteria	Test Result
Negative	Flu A NEGATIVE;Flu B NEGATIVE;RSV NEGATIVE	Passed
Low Positive	Flu A POSITIVE;Flu B POSITIVE;RSV POSITIVE	Passed
High Positive	Flu A POSITIVE;Flu B POSITIVE;RSV POSITIVE	Passed

■ If checked this document is produced electronically and valid without a wet signature.

ThuyTien Nauyen

Oct 16, 2024

Signature of Quality Assurance,

Date

Name: ThuyTien Nguyen

Title: Quality Systems Specialist

XPRSFLURSV-CN-10 1001452780

Final Audit Report 2024-10-16

Created: 2024-10-16

By: ThuyTien Nguyen (thuytien.nguyen@cepheid.com)

Status: Signed

Transaction ID: CBJCHBCAABAABMS4JGp_0uLm4pxq31xb3vmE6xLJ9sN7

"XPRSFLURSV-CN-10 1001452780" History

Document created by ThuyTien Nguyen (thuytien.nguyen@cepheid.com) 2024-10-16 - 4:22:05 PM GMT- IP address: 45.62.187.253

Document emailed to ThuyTien Nguyen (thuytien.nguyen@cepheid.com) for signature 2024-10-16 - 4:22:45 PM GMT

Document e-signed by ThuyTien Nguyen (thuytien.nguyen@cepheid.com)
Signature Date: 2024-10-16 - 4:22:55 PM GMT - Time Source: server- IP address: 45.62.187.253

Agreement completed. 2024-10-16 - 4:22:55 PM GMT