

Title: Quality Systems Specialist

CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements and ISO 13485.

Drug Administration's Quali	ity System Requirements and ISO 13485.		
Product Name: Xpert [®] Xpress CoV-2/Flu/RSV <i>plus</i>		Instructions for Use (IFU)	
Catalogue Part No.: XPRS4PLEX-10		Part No. and Rev.: 302-8057 Rev.	A 2
Kit Lot No.: 1001463045		302-8055 Rev.	С
Cartridge Lot No.: 64006			
Kit Expiration Date: 2025	-12-07		
Legal Manufacturer	Manufacturing Facility Cepheid	Solna	Sunnyvale
Cepheid	121 N Guild Avenue	Lodi	
904 Caribbean Drive		Loui	
Sunnyvale, CA 94089 USA	Lodi, CA 95240 USA		
Functional Testing accordi			
Test Description	Acceptance Criteria		Test Result
,	SARS-CoV-2 NEGATIVE;Flu A NEGATIVE;Flu B		Passed
Negative		NEGATIVE;RSV NEGATIVE	
Positive	SARS-CoV-2 POSITIVE;Flu A POSITIVE;Flu B		Passed
	POSITIVE;RSV POSITI	POSITIVE;RSV POSITIVE	
☐ If checked this documer	nt is produced electronically and valid witho	ut a wet signature.	
Wark Jan	ale 18 DEC 2024		
Signature of Quality Assur	Date Date		
Name: Mark Gonzalez			