



## CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Medical Devices Regulations (CMDR).

**Product Name:** Xpert® MRSA NXG

**Cepheid Catalogue Part No.:** GXMRSA-NXG-10

**Kit Lot No.:** 1001446502

**Cartridge Lot No.:** 22008

**Kit Expiration Date:** 2026-08-30

**Legal Manufacturer**

Cepheid  
904 Caribbean Drive  
Sunnyvale, CA 94089 USA

**Manufacturing Facility**

Cepheid  
904 Caribbean Drive  
Sunnyvale, CA 94089  
USA

- Solna       Sunnyvale  
  
 Lodi

***Functional Testing according to D23916, Rev. U.1***

<b><i>Test Description</i></b>	<b><i>Acceptance Criteria</i></b>	<b><i>Test Result</i></b>
LOW Positive	MRSA detected	Passed
HIGH Positive	MRSA detected	Passed
Negative	MRSA not detected	Passed

If checked, this document is produced electronically and therefore valid without a wet signature

Zenny Santos  
Zenny Santos (Sep 24, 2024 23:38 PDT)

Sep 24, 2024

Signature of Quality Assurance,

Date

**Name:** Zenny Santos

**Title:** Quality Systems Specialist