

CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive, the Canadian Medical Devices Regulations (CMDR), and the China National Medical Products Administration's Quality System Requirements.

Product Name: Xpert® MTI	B/RIF Ultra	
Cepheid Catalogue Part No.	: GXMTB/RIF-ULTRA-50 ▼	
Kit Lot No.: 1001452737		
Cartridge Lot No.: 60701		
Kit Expiration Date: 2026-0	4-12	
Legal Manufacturer	Manufacturing Facility	
Cepheid AB	Cepheid	Solna Sunnyval
Röntgenvägen 5	121 N Guild Avenue	o Soma o Sumy van
SE-17154 Solna	Lodi, CA 95240	
Sweden	USA	Lodi
2001 2 Coccipion	MTB DETECTED VERY LOW; Rif Resistance NOT DETECTED or	1000 1000
Test Description	Acceptance Criteria	Test Result
	MTB DETECTED VERY LOW; Rif Resistance NOT DETECTED or MTB DETECTED LOW; Rif Resistance NOT DETECTED	
Wild Type Control	MTB DETECTED LOW; RIT RESIstance NOT DETECTED MTB DETECTED MEDIUM; Rif Resistance NOT DETECTED	Passed
	or MTB DETECTED HIGH; Rif Resistance NOT DETECTED	
A.	MTB DETECTED VERY LOW;RIF Resistance DETECTED or	
Mutant Control	MTB DETECTED LOW;RIF Resistance DETECTED or MTB DETECTED MEDIUM;RIF Resistance DETECTED	Passed
	MTB DETECTED HIGH;RIF Resistance DETECTED	
Negative	MTB NOT DETECTED	Passed
If checked this document	is produced electronically and therefore valid w	vithout a wet signature
in checked, this document		ittiout a wet signature
	11/1/24	
Signature of Quality Assura	ance, Date	
Name: Mai Kue		

Quality System Specialists

Title: