

Title: Quality Assurance Specialist I

CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and

	red by Cepheid's Quality System, in y System Requirements and ISO 13485.	compliance with	the US Food and
Product Name: Xpert® Xpress CoV-2/Flu/RSV plus		Instructions for Use (IFU) Part No. and Rev.:	
Catalogue Part No.: XPRS4PLEX-10 302-8057 R		302-8057 ▼ Rev	A
Kit Lot No.: 1001456648			
Cartridge Lot No.: 63312			
Kit Expiration Date: 2025-1	1-02		
Legal Manufacturer	Manufacturing Facility	Solna	Sunnyvale
Cepheid	Cepheid	<u> </u>	O Duning vanc
904 Caribbean Drive	121 N Guild Avenue	O Lodi	
Sunnyvale, CA	Lodi, CA 95240 USA		
94089 USA Functional Testing according	g to D47377, Rev. L		T I I
Test Description	Acceptance Criteria		Test Result
Negative	SARS-CoV-2 NEGATIVE;Flu A NEGATIVE;Flu B NEGATIVE;RSV NEGATIVE		Passed
Positive	SARS-CoV-2 POSITIVE;Flu A POSITIVE;Flu B POSITIVE;RSV POSITIVE		Passed
☐ If checked this document	is produced electronically and valid with	out a wet signature.	
anos Sho	10 NOV 20	14	
Signature of Quality Assura	nnce, Date		
N. Anna Cl			
Name: Aaron Shaw			