

## **CERTIFICATE OF ANALYSIS**

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive, the Canadian Medical Devices Regulations (CMDR), and the China National Medical Products Administration's Quality System Requirements.

Product Name: Xpert® Xpress Flu/RSV

Cepheid Catalogue Part No.: XPRSFLU/RSV-CN-10

Kit Lot No.: 1001452781

Cartridge Lot No.: 46508

**Kit Expiration Date: 2026-09-27** 

Legal Manufacturer Manufacturing Facility OSolna

Cepheid Cepheid 904 Caribbean Drive 904 Caribbean Drive Sunnyvale, CA 94089 Sunnyvale, CA 94089

USA USA

Functional Testing

Test Description	Acceptance Criteria	Test Result
Negative	Flu A NEGATIVE;Flu B NEGATIVE;RSV NEGATIVE	Passed
Low Positive	Flu A POSITIVE;Flu B POSITIVE;RSV POSITIVE	Passed
High Positive	Flu A POSITIVE;Flu B POSITIVE;RSV POSITIVE	Passed

■ If checked this document is produced electronically and valid without a wet signature.

Zenny Santos (Oct 16, 2024 14:24 PDT)

Oct 16, 2024

Signature of Quality Assurance, Date

Name: Zenny Santos

Title: Quality Systems Specialist

Sunnyvale