



CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Medical Devices Regulations (CMDR).

Product Name: Xpert® MRSA/SA Blood Culture

Cepheid Catalogue Part No.: GXMRSA/SA-BC-10

Kit Lot No.: 1001448225

Cartridge Lot No.: 22003

Kit Expiration Date: 2026-03-29

Legal Manufacturer

Cepheid
904 Caribbean Drive
Sunnyvale, CA 94089 USA

Manufacturing Facility

Cepheid
904 Caribbean Drive
Sunnyvale, CA 94089
USA



Solna



Sunnyvale



Lodi

Functional Testing according to D36872, Rev. H

<i>Test Description</i>	<i>Acceptance Criteria</i>	<i>Test Result</i>
Positive	MRSA Positive; SA Positive	Passed
Negative	MRSA Negative; SA Negative	Passed

If checked, this document is produced electronically and therefore valid without a wet signature

ThuyTien Nguyen
ThuyTien.Nguyen (Sep 27, 2024 13:17 PDT)

Sep 27, 2024

Signature of Quality Assurance,

Date

Name: ThuyTien Nguyen

Title: Quality Systems Specialist





1001448225

Final Audit Report

2024-09-27

Created:	2024-09-27
By:	ThuyTien Nguyen (thuytien.nguyen@cepheid.com)
Status:	Signed
Transaction ID:	CBJCHBCAABAAAsrBBBU5ROQj9dbwQZE7clyVDz4KO0pa

"1001448225" History

-  Document created by ThuyTien Nguyen (thuytien.nguyen@cepheid.com)
2024-09-27 - 8:15:34 PM GMT- IP address: 45.62.187.255
-  Document emailed to ThuyTien Nguyen (thuytien.nguyen@cepheid.com) for signature
2024-09-27 - 8:16:21 PM GMT
-  Document e-signed by ThuyTien Nguyen (thuytien.nguyen@cepheid.com)
Signature Date: 2024-09-27 - 8:17:03 PM GMT - Time Source: server- IP address: 45.62.187.255
-  Agreement completed.
2024-09-27 - 8:17:03 PM GMT