

CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements and ISO 13485.


Product Name: Xpert® Mpx
Cepheid Catalog Part No.: GXMPX-10
Kit Lot No.: 1001445082
Cartridge Lot No.: 01001
Kit Expiration Date: 2025-08-17

Legal Manufacturer
Cepheid
904 Caribbean Drive
Sunnyvale, CA 94089 USA

Manufacturing Facility
Cepheid
904 Caribbean Drive
Sunnyvale, CA 94089
USA

Functional Testing

<i>Test Description</i>	<i>Acceptance Criteria</i>	<i>Test Result</i>
Positive	MPXV Clade II Detected	Passed
Negative	MPXV Clade II Not Detected; OPXV Not Detected	Passed



Signature of Quality Assurance,



Date

Name: **Molly Doan**

Title: **Quality Systems Specialist**