

CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Devices Regulations.

| Product Name: | Xpert® Xpress CoV-2/Flu/RSV plus |
|----------------------|----------------------------------|

Cepheid Catalogue Part No.: XP3COV2/FLU/RSV-10

Kit Lot No.: 1000361897

Cartridge Lot No.: 05122

Kit Expiration Date: 2022-12-11

Legal ManufacturerManufacturing FacilityOsolnaSunnyvaleCepheidCepheidNewarkLodi IVD (B2)

904 Caribbean Drive 904 Caribbean Drive

Sunnyvale, CA 94089 Sunnyvale, CA 94089

USA USA

Functional Testing

| Test Description | Acceptance Criteria | Test Result |
|------------------|---|-------------|
| Negative | SARS-CoV-2 NEGATIVE;Flu A NEGATIVE;Flu B NEGATIVE;RSV NEGATIVE | Passed |
| Positive | SARS-CoV-2 POSITIVE;Flu A POSITIVE;Flu B POSITIVE;RSV POSITIVE | Passed |

If checked this document is produced electronically and valid without a wet signature.

Signature of Quality Assurance, Date

Name: Jenny Nguyen

Title: Quality Assurance Supervisor