

CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Devices Regulations.

Product Name: Xper	rt® Xpress CoV-2/Flu/RSV plus		
Cepheid Catalogue Part	No.: XP3COV2/FLU/RSV-	10	
Kit Lot No.: 1000364486			
Cartridge Lot No.: 05206			
Kit Expiration Date: 2022	2-12-18		
Legal Manufacturer	Manufacturing Facility	Solna	Sunnyvale
Cepheid	Cepheid	Newark	Lodi IVD (B2)
904 Caribbean Drive	904 Caribbean Drive		

Functional Testing

USA

Sunnyvale, CA 94089

Test Description	Acceptance Criteria	Test Result
Negative	SARS-CoV-2 NEGATIVE;Flu A NEGATIVE;Flu B NEGATIVE;RSV NEGATIVE	Passed
Positive	SARS-CoV-2 POSITIVE;Flu A POSITIVE;Flu B POSITIVE;RSV POSITIVE	Passed

If checked this document is produced electronically and valid without a wet signature. 07/02/2022

Sunnyvale, CA 94089

USA

Signature of Quality Assurance,

Date

Name: Jenny Nguyen

Title: Quality Assurance Supervisor