



CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Medical Device Regulations (CMDR).

Product Name: Xpert® MRSA

Cepheid Catalogue Part No.:

Kit Lot No.:

Cartridge Lot No.:

Kit Expiration Date:

Legal Manufacturer

Cepheid
904 Caribbean Drive
Sunnyvale, CA 94089 USA

Manufacturing Facility

Solna Sunnyvale

Cepheid
904 Caribbean Drive
Sunnyvale, CA 94089 USA

Functional Testing

<i>Test Description</i>	<i>Acceptance Criteria</i>	<i>Test Result</i>
Positive	MRSA Positive	Passed
Negative	MRSA Negative	Passed

23 FEB 23

Signature of Quality Assurance

Date

Name:

Title: