

CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Medical Device Regulations (CMDR).

Product Name:_ Xpert MTB/XDR			
Cepheid Catalogue Part No.: GXMTB/XDR-10	to the state of th		
Kit Lot No.: 1001409456			
Cartridge Lot No.: 03806			
Kit Expiration Date: 2025-10-05		٠	
Legal Manufacturer	Manufacturing Facility	○ Solna	Sunnyvale
Cepheid AB	Cepheid		
Röntgenvägen 5	904 Caribbean Drive		
SE-17154 Solna	Sunnyvale, CA 94089 USA		
Sweden			

Functional Testing

Test Description	Acceptance Criteria	Test Result
Negative	MTB NOT DETECTED	Passed
Positive - Wild Type	MTB DETECTED;INH Resistance NOT DETECTED;FLQ Resistance NOT DETECTED;AMK Resistance NOT DETECTED;KAN Resistance NOT DETECTED;CAP Resistance NOT DETECTED;ETH Resistance NOT DETECTED	Passed
Positive - Mutant	MTB DETECTED;INH Resistance DETECTED;FLQ Resistance DETECTED;AMK Resistance DETECTED;KAN Resistance DETECTED;CAP Resistance DETECTED;ETH Resistance DETECTED	Passed

Thurstun 01	116/2024
Signature of Quality Assurance	Date
Name: ThuyTien Nguyen	
Title: Quality Systems Specialist	