

## **CERTIFICATE OF ANALYSIS**

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Medical Devices Regulations (CMDR).

Product Name: Xpert® TV			
Cepheid Catalogue Part No.: GXTV-1	0		
Kit Lot No.: 1001461351			
Cartridge Lot No.: 22202			
Kit Expiration Date: 2026-06-07			
Legal Manufacturer	Manufacturing Facility		
Cepheid	Cepheid	Solna	Sunnyvale
904 Caribbean Drive	121 N Guild Avenue	Donia (	Dunny van
Sunnyvale, CA 94089 USA	Lodi, CA 95240		
	USA	<ul><li>Lodi</li></ul>	

Functional Testing according to D20618, Rev. M

Test Description	Acceptance Criteria	Test Result
LOW Positive	TV detected	Passed
HIGH Positive	TV detected	Passed
Negative	TV not detected	Passed

If checked, this document is produced electronically and therefore valid without a wet signature

Signature of Quality Assurance, Date

Name: Betty Thao

Title: Quality Systems Specialist