



CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the Japanese Pharmaceuticals and Medical Devices Law and ISO 13485.

Product Name: Xpert® CT/NG

Cepheid Catalogue Part No.: GXCT/NG-JP-10

Kit Lot No.: 1001458710

Cartridge Lot No.: 36205

Kit Expiration Date: 2026-10-11

Legal Manufacturer

Cepheid AB
Röntgenvägen 5
SE-171 54 Solna,
Sweden

Manufacturing Facility

Cepheid AB
Röntgenvägen 5
SE-171 54 Solna,
Sweden

Functional Testing

<i>Test Description</i>	<i>Acceptance Criteria</i>	<i>Test Result</i>
LOW Positive	CT DETECTED;NG DETECTED	Passed
MEDIUM Positive	CT DETECTED;NG DETECTED	Passed
HIGH Positive	CT DETECTED;NG DETECTED	Passed
Negative	CT NOT DETECTED; NG NOT DETECTED	Passed

Final Judgement

<i>Product Specification</i>	<i>Judgement</i>
Sensitivity	Passed
Precision	Passed
Simultaneous Repeatability	Passed

If checked this document is produced electronically and valid without a wet signature.

Lava Maroof 20241119

Signature of Quality Assurance

Date

Name: Lava Maroof

Title: QA Analyst