

CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Devices Regulations.

| Product Name: Xper | rt® Xpress CoV-2/Flu/RSV plus | | |
|---|--|------------------|---------------------------|
| Cepheid Catalogue Part | No.: XP3COV2/FLU/RSV-10 | | |
| Kit Lot No.: 1001457644 Cartridge Lot No.: 63607 Kit Expiration Date: 2025 | 5 11 16 | | |
| Legal Manufacturer Cepheid 904 Caribbean Drive Sunnyvale, CA 94089 USA Functional Testing | Manufacturing Facility Cepheid 121 N Guild Avenue Lodi, CA 95240 USA | OSolna Newark | Sunnyvale • Lodi IVD (B2) |

| Test Description | cription Acceptance Criteria | |
|------------------|---|--------|
| Negative | SARS-CoV-2 NEGATIVE;Flu A NEGATIVE;Flu B NEGATIVE;RSV NEGATIVE | Passed |
| Positive | SARS-CoV-2 POSITIVE;Flu A POSITIVE;Flu B POSITIVE;RSV POSITIVE | |

☐ If checked this document is produced electronically and valid without a wet signature.

| | US . | 11/24/24 |
|--------|---------------------------|----------|
| Signat | ure of Quality Assurance, | Date |
| Nama | Asia Taylor Torres | |

Title: Quality Systems Specialist