

CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Medical Devices Regulations (CMDR).

Product Name: Xpert® HPV

Catalogue Part No.: GXHPV-CE-10

Kit Lot No.: 1001460185

Cartridge Lot No.: 30904

Kit Expiration Date: 2026-05-17

Legal Manufacturer

Cepheid AB Röntgenvägen 5 SE-171 54 Solna,

Sweden

Manufacturing Facility

Cepheid AB Röntgenvägen 5 SE-171 54 Solna,

Sweden

Functional Testing

Test Description	Acceptance Criteria	Test Result
LOW Positive	HPV16 Positive, HPV18/45 Positive, other high risk HPV Positive	Passed
HIGH Positive	HPV16 Positive, HPV18/45 Positive, other high risk HPV Positive	Passed
Negative	HPV16 Negative, HPV18/45 Negative, other high risk HPV Negative	Passed

If checked this document is produced electronically and valid without a wet signature.

Signature of Quality Assurance	Date
Robert Fiedler Robert Fiedler (Nov 27, 2024 11:56 GMT+1)	

Name: Robert Fiedler

Title: QA Analyst